

PCT

WORLD INTELLECTUAL PROPERTY ORGANIZATION
International Bureau



57E/L/2/62


INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(51) International Patent Classification ⁶ : G01S 7/52		(11) International Publication Number: WO 98/20361
A1		(43) International Publication Date: 14 May 1998 (14.05.98)
(21) International Application Number: PCT/US97/20162 (22) International Filing Date: 4 November 1997 (04.11.97) (30) Priority Data: 08/746,360 8 November 1996 (08.11.96) US (71) Applicant: RESEARCH CORPORATION TECHNOLOGIES, INC. [US/US]; Suite 600, 101 N. Wilmot Road, Tucson, AZ 85711-3335 (US). (72) Inventor: CHRISTOPHER, Ted; Apartment 1, 51 Lilac Drive, Rochester, NY 14620 (US). (74) Agents: DiGIGLIO, Frank, S. et al.; Scully, Scott, Murphy & Presser, 400 Garden City Plaza, Garden City, NY 11530 (US).		(81) Designated States: AU, CA, JP, European patent (AT, BE, CH, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE). Published <i>With international search report. Before the expiration of the time limit for amending the claims and to be republished in the event of the receipt of amendments.</i>

(54) Title: FINITE AMPLITUDE DISTORTION-BASED INHOMOGENEOUS PULSE ECHO ULTRASONIC IMAGING

(57) Abstract

A method and system for imaging a sample. The method includes the steps of generating an ultrasonic signal, directing the signal into a sample, which signal is distorted and contains a first order and higher order component signals at first and higher frequencies respectively. The received distorted signal is processed, and an image is formed, and then displayed, from one of the higher order component signals of the received distorted signal.



FOR THE PURPOSES OF INFORMATION ONLY

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AL	Albania	ES	Spain	LS	Lesotho	SI	Slovenia
AM	Armenia	FI	Finland	LT	Lithuania	SK	Slovakia
AT	Austria	FR	France	LU	Luxembourg	SN	Senegal
AU	Australia	GA	Gabon	LV	Latvia	SZ	Swaziland
AZ	Azerbaijan	GB	United Kingdom	MC	Monaco	TD	Chad
BA	Bosnia and Herzegovina	GE	Georgia	MD	Republic of Moldova	TG	Togo
BB	Barbados	GH	Ghana	MG	Madagascar	TJ	Tajikistan
BE	Belgium	GN	Guinea	MK	The former Yugoslav Republic of Macedonia	TM	Turkmenistan
BF	Burkina Faso	GR	Greece	ML	Mali	TR	Turkey
BG	Bulgaria	HU	Hungary	MN	Mongolia	TT	Trinidad and Tobago
BJ	Benin	IE	Ireland	MR	Mauritania	UA	Ukraine
BR	Brazil	IL	Israel	MW	Malawi	UG	Uganda
BY	Belarus	IS	Iceland	MX	Mexico	US	United States of America
CA	Canada	IT	Italy	NE	Niger	UZ	Uzbekistan
CF	Central African Republic	JP	Japan	NL	Netherlands	VN	Viet Nam
CG	Congo	KE	Kenya	NO	Norway	YU	Yugoslavia
CH	Switzerland	KG	Kyrgyzstan	NZ	New Zealand	ZW	Zimbabwe
CI	Côte d'Ivoire	KP	Democratic People's Republic of Korea	PL	Poland		
CM	Cameroon	KR	Republic of Korea	PT	Portugal		
CN	China	KZ	Kazakhstan	RO	Romania		
CU	Cuba	LC	Saint Lucia	RU	Russian Federation		
CZ	Czech Republic	LI	Liechtenstein	SD	Sudan		
DE	Germany	LK	Sri Lanka	SE	Sweden		
DK	Denmark	LR	Liberia	SG	Singapore		
EE	Estonia						

-1-

1

FINITE AMPLITUDE DISTORTION-BASED
INHOMOGENEOUS PULSE ECHO ULTRASONIC IMAGING

5

This invention generally relates to ultrasonic pulse echo imaging, and more specifically, to ultrasonic pulse echo imaging based on the distortion of ultrasonic signals transmitted into samples.

10

Ultrasonic pulse echo imaging is widely used in many medical applications. While this technique has received wide acceptance, it would be desirable to improve the resolution of the images formed from this technique. For example, ultrasonic pulse echo imaging in inhomogeneous media suffers from significant lateral and contrast resolution losses due to the defocusing effects of the inhomogeneities of the media. The losses in lateral and contrast resolution are associated with increases in the width of the main beams and increases in side lobe levels, respectively.

15

20

25

30

These two forms of resolution loss represent a significant hurdle to improving the clinical utility of biomedical ultrasonic imaging. A number of research efforts are currently underway to investigate the defocusing effects of tissue and to consider corrective measures. These efforts, however, generally assume linear propagation and base the image-formation process on the reception of the transmitted pulse.

35

-2-

The present invention is directed to
1 improved ultrasonic imaging and methods, as well as
improved resolution of images formed from ultrasonic
echo signals. The present invention utilizes the
enhanced inhomogeneous focusing properties of the
5 finite amplitude distortion generated higher harmonics
of an ultrasonic imaging beam in order to obtain
improved contrast resolution and lateral resolution
images.

More specifically, the present invention is
10 directed to a method which includes the steps of
generating an ultrasonic signal, directing the signal
into a sample, which signal is distorted and contains
a first order and higher order component signals at
first and higher frequencies respectively. The
15 received distorted signal is processed, and an image
is formed, and then displayed, from one of the higher
order component signals of the received distorted
signal.

With the preferred embodiment of the
20 invention disclosed herein in detail, the ultrasonic
image is based on one of the received finite amplitude
distortion component (or nonlinearly-generated higher
harmonics) associated with the transmitted signal. In
the simplest case, in which the transducer emits
25 negligible energy in the second harmonic bandwidth,
such an image can be formed by adding an initial high
pass filtering of the received signal. In general,
such an image can be formed by using a two pulse
transmit, receive, normalize, and then high pass
30 filtering scheme. Such a two pulse scheme can be used

-3-

1 to remove the source or linear content in the second
and higher harmonic bandwidths.

Further benefits and advantages of the invention will become apparent from a consideration of the following detailed description given with
5 reference to the accompanying drawings, which specify and show preferred embodiments of the invention.

Figure 1 shows an ultrasonic imaging system embodying the present invention.

10 Figures 2a, 2b and 2c show various parameters associated with the linear propagation results for a focused 2 MHz Gaussian Transducer operating in a liver medium.

Figure 3 shows discrete harmonic velocities used to compute various values associated with a 2 MHz
15 propagation in a liver medium.

Figures 4a and 4b display nonlinear propagation results for a focused 2 MHz Gaussian transducer.

20 Figure 5a shows the log-scaled, normalized one-way focal plane profiles of the 2 MHz fundamental, the 4 MHz second harmonic, and the 4 MHz fundamental, and Figure 5b shows the corresponding two way profiles for these beam patterns.

25 Figure 6a shows the on-axis source plane and the subsequent focal plane for a 2 MHz Gaussian source.

Figure 6b illustrates the corresponding spectrums of the source and focal planes shown in Figure 6a.

30 Figure 6c shows the focal nonlinear

35

-4-

distortion pulse obtained by constructing the waveform
1 using only the focal spectral information shown in
Figure 6b from 3 to 8 MHz.

Figure 7 shows the focal plane profile from
a 2 MHz continuous wave propagation and a 2 MHz
5 centered pulse propagation.

Figure 8 is a table giving the on-axis
source plane intensity versus the received second and
third harmonic levels and focal field parameters.

Figure 9 is a Table providing focal and
10 received second harmonic levels versus focal length.

Figures 10a and 10b show the log-scaled
first, second and third harmonic axial amplitudes for
the focused 2 MHz Gaussian transducer and the
corresponding log-scaled focal plane radial beam
15 profiles.

Figure 11 illustrates scaled axial
amplitudes for a 4 MHz second harmonic and 4 MHz
fundamental harmonic beams in a liver medium.

Figures 12a-12f depict pairs of one-way
20 focal plane harmonic amplitude diameters and the
corresponding average radii for an unjittered, or
homogeneous, path and an abdominal wall-jittered
propagation path.

Figures 13a-13d show two-way average radii
25 results for the abdominal wall-jittered propagation
path represented in Figures 12a-12f.

Figures 14a and 14b display normalized two-
way averaged radial results from five abdominal wall-
jittered propagations for the 2 MHz fundamentals, the
30 4 MHz second harmonics and the 4 MHz fundamentals, and

the corresponding radially integrated magnitudes.

1 Figures 15a and 15b show normalized two-way
averaged radial results from five abdominal wall-
jittered propagations for the 4 MHz fundamentals, the
8 MHz second harmonics and the 8 MHz fundamentals, and
5 the corresponding radially integrated normalized
magnitudes.

 Figures 16a and 16b show normalized two-way
averaged radial results from five breast jittered
propagations for the 2 MHz fundamentals, the 4 MHz
10 second harmonics and the 4 MHz fundamentals, and the
corresponding radially integrated magnitudes.

 Figure 17 is a table giving the -20dB full-
widths for the average two-way profiles shown in
Figures 14a, 15a and 16a and the full-widths at the
15 0.9 level for the integrated profiles of Figures 14b,
15b and 16b. Also shown are the corresponding results
from the 8 MHz breast jittered propagations.

 Figures 18a-18d show an imperfect source
pulse, linearly-scaled and log-scaled focal spectrums
20 of that pulse, and the corresponding nonlinear
distortion pulse obtained by constructing with the
spectral information in Figure 18c.

 Figure 19a shows the log-scaled focal
spectrum of Figure 19c overlaid with the focal
25

spectrum from the same source using a half amplitude
version of the source pulse, as depicted in Figure
19c.

30 Figure 19b shows the resulting difference

-6-

1 spectrum computed for the two spectrum shown in Figure
20a.

Figure 19c shows the corresponding nonlinear
distortion pulse obtained by constructing with the
spectral information in Figure 19b and starting at
5 2.75 MHz.

Figure 1 illustrates ultrasonic imaging
system 10. A pulse generator 12 and a function
generator 14 produce a sinusoidal pulse ultrasonic
10 signal of, for example, 2.0 MHz at a pulse repeat
frequency of, for instance, 1 kHz. This signal is
sent to amplifier 20, which amplifies the signal and
transmits the amplified signal to transducer-receiver
unit 22, and this unit then transmits the signal into
15 sample 24.

In this sample 24, the input signal is both
distorted and reflected. The distortion creates a
distorted signal having a multitude of component
signals, each of which has a respective frequency or
20 frequency bandwidth. The distorted signal is
reflected by sample 24, and this reflected signal is
received by transducer-receiver unit 22, amplified by
pre-amplifier 30, and then further amplified by
amplifier 32. The received and amplified signal is
25 then sent through a high-pass filter 34 to enhance the
relative strength of the desired higher harmonic
component of the received signal. The resulting
signal is digitized in analog-to-digital converter 36,
and then processed by processor 40 to produce an
30 image.

35

-7-

This image may be displayed on a video monitor
1 42, stored on a video cassette recorder (VCR) 44,
output on a printer device 46, or stored in any of a
variety of hard copy storage devices 50, such as
5 medical film recorders, digital tape machines, optical
disks, magnetic tapes and disks or the like. Suitable
means may be used to move the focal point of the
transmitted signal around sample 24. For instance,
transducer-receiver unit 22 may be a phased array unit
having electrical circuitry to move the focal point of
10 the transmitted signal. Alternatively, a motor 52 may
be employed to move transducer-receiver unit 22 and
thereby move the focal point of the transmitted signal
around sample 24.

System 10 can also be operated in a two
15 pulse scheme or mode. In this mode of operation,
system 10 generates and transmits into sample 24 two
different pulses. Preferably, the transmitted signals
are identical except that one of them is scaled up in
pressure. The pulses are transmitted one after
20 another with, for example, approximately 1/4000 second
interval between them. The reflected, distorted
signals from both pulses are received by transducer
unit 22, and these signals are then digitized in
analog-to-digital converter 36. The digital data
25 values obtained from the first pulse are stored,
scaled and then subtracted from the digital data
values obtained from the second pulse, producing a
difference or resultant signal. Subsequent processing
of this difference signal in system 10 is the same as
30 in the above-described one pulse case. One suitable

- 8 -

1 procedure for scaling the digital data values obtained
from the first pulse is discussed below.

Several computational models exist which can accurately describe the finite amplitude propagation of a continuous beam. Such models can be extended to
5 compute the case of a propagating pulse as well. These models account for the effects of diffraction, nonlinearity and absorption. One of these models, referred to as the NLP model is described in "New approaches to nonlinear diffractive field propagation"
10 J. Acoust. Soc. Am. 90, 488-499 (1991) by P.T. Christopher and K.J. Parker, the disclosure of which is herein incorporated by reference, and this model was used to compute the following linear and nonlinear examples. This model has been updated to account for
15 the effects of dispersion. The associated harmonic velocities were computed using an algorithm disclosed in "Modeling acoustic field propagation for medical devices," Ph.D. thesis, U. of Rochester (1993) by T. Christopher, the disclosure of which is herein
20 incorporated by reference.

The propagations discussed below are for a Gaussian apodized, axially symmetric focused source. This form of device offers excellent image quality and produces a field comparable to that of the two
25 dimensional array-based transducers now being developed.

Many biomedical imaging devices are not axially symmetric, though. The finite amplitude beams produced by such devices are well described by the
30 computations for the axially symmetric transducers

35

-9-

discussed here in detail, though. In measuring the
1 nonlinear harmonic generation from an unfocused
rectangular source, Kamakura, Tani, Kumamoto and Ueda
noted in "Harmonic generation in finite amplitude
sound beams from a rectangular aperture source," J.
5 Acoust. Soc. Am. 97, 3510-3517 (1995), that "the
[nonlinear] harmonic pressure levels in the far field
[were] almost the same as from a circular aperture
source with equal face area and equal initial
pressure, independent of the source levels." Though
10 this result was obtained for only one device (with a
ratio of source side lengths or aspect ratio of 11 to
6), the higher harmonic pressure levels associated
with a non-axially symmetric device are approximately
equal to those of the corresponding axially symmetric
15 source. More importantly, the lack of axial symmetry
does not affect the relative sidelobe advantages
exhibited by the nonlinearly-generated harmonics in a
homogeneous propagation. These homogeneous path
sidelobe level advantages are the basis for the
20 imaging-relevant advantages of the higher harmonics in
an inhomogeneous propagation.

Figures 2a, 2b and 2c show various
parameters associated with the linear propagation
results for a focused 2 MHz Gaussian transducer
25 operating in a liver medium. In particular, the
source plane amplitude profile, the on-axis amplitude,
and the radial focal plane (at $Z = 6\text{cm}$) beam profile
are shown in Figures 2a, 2b and 2c respectively. The
results shown in Figures 2a, 2b and 2c were obtained
30 by computing the linear, liver path propagation of the

-10-

field of a focused 2 MHz Gaussian source using the NLP beam propagation model. The NLP model propagates a planar, normal velocity description of the acoustic field. No inhomogeneities or phase aberrations were accounted for in this propagation or any of the subsequent propagations considered immediately below. The relevant liver propagation parameters used were $c = 1570$ m/s, $\rho = 1.05$ g/cm³, $\alpha = 0.03$ Np/cm and $b = 1.3$ (where α and b are the coefficients describing absorption in a power law form).

The Gaussian shading of the magnitude of the source plane normal velocity field was such that the half-amplitude radial distance was 0.84 cm. The on-axis, source plane RMS acoustic intensity ($\rho c |u|^2 / 2$, where u is the acoustic particle velocity) for the field was 2 W/cm². The radial extent of the source was 1.5 cm. The source plane field was focused using a spherically-focusing factor ($e^{j\theta(r)}$, where $\theta(r) = (2\pi f/c)\sqrt{r^2 + F^2}$). The geometric focal length F was 6 cm and the sound speed (c) used to compute $\theta(r)$ was that of water (1500 m/s).

Figures 2a and 2b depict the normal velocity magnitudes of the Gaussian transducer's source plane and on-axis fields respectively. Figure 2c displays the focal plane ($z = 6$ cm) radial profile of the 2 MHz field. The drop in the magnitude of the field from the mainlobe to the first sidelobe in Figure 2c is 36 dB. In the absence of strong medium phase aberration this allows the device to produce high contrast images.

The same 2 MHz Gaussian source was then propagated nonlinearly through the same liver path.

-11-

The nonlinear parameter β used to represent liver was 4.7. The NLP model used 4 harmonics (2, 4, 6 and 8 MHz) to compute the pre-focal region ($z = 0$ to $z = 3$ cm) propagation and up to 10 harmonics to represent the subsequent focal and post-focal region propagation. The harmonic velocities were computed using the algorithm described in "Modeling acoustic field propagation for medical devices," T. Christopher, Ph.D. Thesis, University of Rochester (1993). The fundamental or 2 MHz component had a propagation speed of 0.157 cm/microsecond (given above as c). The discrete harmonic velocities used by NLP to compute the 2 MHz propagation are shown in Figure 3.

Figure 4a displays the axial magnitudes of the fundamental, second harmonic, and third harmonic fields, at 102a, 102b and 102c respectively, as computed for the nonlinear propagation. The fundamental or 2 MHz axial curve is only slightly different from the corresponding linear curve shown in Figure 3b. At $z = 6$ cm the nonlinear 2 MHz curve is about 1% lower than the 2 MHz linear curve. This difference was due to growth of the higher harmonics in the nonlinear propagation. In Figure 4b the corresponding focal ($z = 6$ cm) pressure waveforms from the linear and nonlinear computations are displayed at 104a and 104b respectively. The pressure waveforms were obtained by converting NLP's normal velocity output to pressure using the impedance relation (all subsequent pressure waveforms were obtained in this way).

The 2 MHz or fundamental beam pattern 104b

-12-

associated with the nonlinear propagation is almost identical with the 2 MHz beam pattern 104a of the linear propagation. Only in a linearly scaled overlay plot of the two beam patterns are there visible differences. These differences are very small and are limited to the near axis portions of the beam patterns. Only at much higher source intensities are the effects of nonlinearity significant to the details of the fundamental's field. These results are consistent with the empirical observation that linear modeling of biomedical ultrasonic device fields accurately describes their (linear-based homogeneous path) imaging performance.

Figure 5a depicts the 2 MHz fundamental and 4 MHz second harmonic focal plane beam amplitude profiles at 106a and 106b. Also shown in Figure 5a at 106c is the corresponding 4 MHz fundamental profile. The 4 MHz fundamental result was obtained by computing the linear propagation of the same Gaussian transducer operating at a source frequency of 4 MHz. All three beam profiles in Figure 5a have been normalized to have on-axis field magnitudes of 1. The finite amplitude distortion-generated second harmonic focal profile 106b has a slightly broader mainlobe than the corresponding (4 MHz) fundamental profile 106c. The radial half-amplitude distance of the second harmonic profile is 36% greater than that of the 4 MHz fundamental profile (0.0983 cm versus 0.0723 cm). The second harmonic profile also has much lower sidelobes than the 4 MHz fundamental profile.

For imaging purposes, the two-way focal plane beam pattern of the Gaussian transducer is of

-13-

interest. The two-way focal beam pattern accounts for both the characteristics of the transmitted pulse in the focal plane and the corresponding sensitivity of the transducer to pulses reflected back from this plane. For linear propagations, the two-way beam pattern for a given depth can be obtained by squaring the corresponding transmit or one-way beam pattern. In Figure 5b the normalized two-way linear beam patterns for the Gaussian transducer operating at 2 and 4 MHz are depicted at 110a and 110b. These curves were obtained by squaring the corresponding one-way or transmit beam patterns shown in Figure 5a.

Also shown in Figure 5b at 110c, is the two-way focal plane beam pattern associated with the 4 MHz second harmonic field. Since the amplitudes of the reflected pulses are much smaller than the transmitted pulses, the propagation of the reflected field back to the transducer is essentially linear. Thus, the two-way focal plane beam pattern shown at 110b for the second harmonic was obtained by multiplying the corresponding one-way pattern 106b shown in Figure 5a by the 4 MHz fundamental one-way pattern 106b also shown in Figure 5a.

The second harmonic's two-way beam pattern has a half-amplitude mainlobe width (or -6 dB beamwidth) that is 12% greater than that of the corresponding beam pattern 106c. The -20 dB beamwidth of the second harmonic is 13% greater than that of the 4 MHz fundamental. The sidelobe advantage displayed in the focal plane profiles of Figure 5a is maintained in the two-way results shown in Figure 5b.

These homogeneous results show that the

-14-

second harmonic field of a focused, apodized transducer offers advantages in contrast resolution
1 over the corresponding linear transducer field.

A pulse propagation was next considered for the 2 MHz Gaussian source. The on-axis, source plane pressure pulse used is displayed in Figure 6a as the
5 desired curve 112a. This pulse was computed by applying a Gaussian window to a 2 MHz cosine. The peak pressure of the pulse was the same as for the 2 W/cm² continuous case considered above. The radial amplitude shading or apodization and the spherical
10 focusing of the source field were also the same as in the previous continuous wave case. The initial source plane pulse consisted of 128 samples across 8 microseconds.

The magnitude of the Fourier transform of
15 the 8 microsecond long source pulse is shown at 114a in Figure 6b. A straightforward implementation of the nonlinear imaging system and method requires negligible overlap between the sources's spectral bandwidth and that of the nonlinear second harmonic
20 (more generally, this would also insure negligible overlap between any of the successive harmonic spectral bands). The source spectrum 114b shown in Figure 6b meets this requirement. Nonlinear images based on source pulses with broader spectrums or with
25 significantly more energy in the second harmonic bandwidth than the one depicted at 112b and 114b in Figures 6a and 6b could be obtained by using an alternative nonlinear imaging scheme described below.

The source plane was then defined using the
30 64 harmonic Fourier transform of the source pulses.

-15-

This multiharmonic source radius was then input into a pulse-propagating version of the NLP model (a model
1 similar to the lithotripter model presented in
"Modeling the Dornier HM3 Lithotripter." T.
Christopher, J. Acoust. Soc. Am. 3088 - 3095 (1994).
The focal output of the resulting nonlinear pulse
5 propagation is also shown in Figures 6a and 6b as
solid curves. The focal pulse waveform has slightly
smaller peak positive and negative pressures than the
corresponding continuous waveform shown in Figure 4b.
Consistent with the smaller amplitudes, the focal
10 pulse is also less distorted than the continuous
waveform. The ratio of the second harmonic's focal
amplitude to that of the fundamental's for this pulse
propagation was 70% of the same ratio for the
corresponding continuous source considered above.

15 In Figure 6c, the waveform associated with
only the spectral bands of the second, third, and the
first half of the fourth harmonic (3 through 8 MHz) is
shown. This waveform was computed as a high pass
filtered reconstruction of the spectral information
20 depicted in Figure 6b. A rectangular window with a
transition at 3 MHz was used in filtering the
transform data. The straightforward nonlinear imaging
approach disclosed herein may use a distortion pulse
like that shown in Figure 6c in order to image the
25 scattering medium.

Not shown for the pulse propagation
considered are the axial and radial harmonic
descriptions. These descriptions were found to be
identical in form to those computed for the nonlinear
30 propagation of the continuous 2 MHz field. All other

-16-

Gaussian-windowed cosine pulses were found to produce identical axial and radial harmonic patterns. The temporal peak amplitude profile of the high pass filtered-focal plane data also very closely followed the form of the 4 MHz second harmonic's amplitude profile. The temporal peak amplitude profile for the pulse propagation considered herein is shown at 116a in Figure 7. This profile was computed using the focal plane frequency data in the range of 3 to 8 MHz (like the waveform in Figure 6c). Also shown at 116b and 116c in Figure 7 are the corresponding 4 MHz second harmonic profile and a second harmonic bandpass filtered (3 to 5 MHz)-temporal peak amplitude profile for the same pulse propagation. The similarity of the two peak amplitude profiles suggests that there is very little energy above 5 MHz for this pulse propagation. The nonlinear harmonic beam and peak amplitude patterns produced by any reasonable source pulse can be computed using an appropriate continuous approximation of the source.

In order for the nonlinearly-generated higher harmonics to be available for imaging in an inhomogeneous media, the received higher harmonics amplitudes cannot be too low relative to the received linear (or transmitted) signal and the transducer's dynamic range. Also, to be available for imaging use, preferably the associated in vivo field amplitudes have a mechanical index less than 1.9.

For the previously considered nonlinear 2 MHz, liver-path propagation, the amplitudes at the focus ($z = 6$ cm) of the fundamental and second harmonic were 0.943 MPa and 0.166 MPa, respectively.

-17-

The second harmonic amplitude was thus 15.09 dB below the amplitude of the 2 MHz transmitted wave. As a wave reflected at the focal point travels the 6 cm back to the transducer/receiver, frequency dependent attenuation reduces the second harmonic by an additional 5.62 dB relative to the 2 MHz component. The received 4 MHz component would thus be 20.71 dB below the transmitted 2 MHz component. The corresponding figures for the 6 MHz third harmonic are 27 dB down at the focus and 39.23 dB down for received signals. This calculation, it should be noted, does not include the effect of the possible additional relative weakening of the second harmonic received signal due to coherent reflectors at the focus in combination with the smaller mainlobe of the second harmonic. This effect is not significant to many bio-ultrasound imaging applications, though.

Table 1 of Figure 8 displays the second and third harmonic received levels at on-axis, source plane intensity values of 0.5, 1, 2, 4 and 8 W/cm². For a given source plane intensity, pulsed devices would have slightly larger received level differentials than those shown in Table 1. For the pulsed propagation considered above, this additional gap would be 0.86 dB for the second harmonic (based on second harmonic bandpass reconstruction and a comparison of the received peak positive pressure levels).

Current biomedical ultrasonic imaging transducers have dynamic ranges of about 100 dB. Even with decreased sensitivity above the transmit frequencies, these devices are capable of creating

-18-

second harmonic images. This capability has been demonstrated by the creation of second harmonic contrast agent-response images. Alternatively, a separate receiver device with appropriate frequency response in the desired nonlinear distortion bandwidth can be used.

5 The effect of focal length on the received second harmonic levels for this Gaussian transducer operating at 2 W/cm² is shown in Table 2 of Figure 9. From a focal length of 4 cm up to a focal length of 12 cm the received second harmonic levels dropped off by 10 7.32 db. Also shown in Table 2 are the corresponding focal second harmonic levels. These levels remained very constant and thus revealed the decreases in received levels as almost entirely due to increased return trip distances.

15 The peak positive and negative pressures of the in vivo nonlinear waveform shown in Figure 4b were 1.12 and -0.84 MPa, respectively. The -0.84 peak rarefaction pressure corresponds to a mechanical index (MI) of 0.59. The highest preferred level for the 20 mechanical index is 1.9. Shown in Table 1 are values of the computed minimum focal pressure and associated mechanical index for this and four other values of source plane intensity. The minimum pressures and thus mechanical indices given in Table 1 have been 25 corrected for the effects of nonlinearity. A linear-only computation would result in larger negative pressures and MI values, in particular at the highest two source intensity levels.

30 The numbers shown in Table 1 show that for in vivo propagations similar to the one considered

- 19 -

here, finite amplitude distortion-based images are readily obtainable within the current mechanical index safety limit. Even at the lowest source intensity case considered (0.5 W/cm^2), a largely second harmonic-based image can be obtained by simply filtering out the transmitted frequency or frequencies. Additionally the second and third harmonic received levels offer some real-time feedback on the magnitude of the focal field amplitudes themselves. Finally, the results displayed in Table 2 suggest that second harmonic imaging may be available at a wide range of focal depths.

The formation of the higher harmonic constituent beams in a propagation finite amplitude beam is a continuous process. In the case of the 2 MHz Gaussian-shaded, focused beam considered above, the 4 MHz second harmonic, the 6 MHz third harmonic, and additional higher harmonics are continuously and cumulatively produced by the beam as it propagates away from the source. Of interest here is the production and focusing of these nonlinear higher harmonic beams between the source and focal plane.

The origin of the higher harmonic beams is the ongoing nonlinear distortion of the propagating waves comprising the (total harmonic) focused beam. The physical effects of diffraction and absorption concurrently act on the higher harmonic beams and thus further define their propagation as well as contribute to changes in the resulting focused beam. The NLP model assumes that the nonlinear or finite amplitude distortion acts in a plane wave fashion on the waves comprising the focused beam. The NLP model uses the

-20-

frequency domain solution to Burgers' equation in an incremental Δz fashion to account for this plane wave distortion approximation.

The frequency domain solution to Burgers' equation as used in the NLP model can be written

$$\frac{\partial u_n}{\partial z} = j \frac{\beta \pi f}{2c^2} \left(\sum_{k=1}^{n-1} k u_k u_{n-k} + \sum_{k=n}^N n u_k u_{k-n}^* \right), \quad n = 1, 2, \dots, N$$

where f is the fundamental frequency and u_n is the n th term in an N term complex Fourier series describing the temporal normal velocity waveform at a radial point in the radial description of the field. The first summation term in the parentheses represents the accretion of the n th harmonic by nonlinear combination of other harmonics that have a sum frequency of nf . The second summation term (with conjugation) represents the depletion of the n th harmonic to other harmonics with a difference frequency of nf . For the case of a Fourier representation of a (periodic) pulse waveform, this accretion and depletion of harmonics results in some interesting phenomenon including the production of a distortion bandwidth below the fundamental bandwidth.

Of interest are the terms in equation (1) contributing to the growth of the second harmonic and, to a lesser extent, the third harmonic. When n equals 2 in equation (1), the positive

contributions to $\frac{\partial u_2}{\partial z}$ come from the $1u_1u_1$ term comprising

-21-

the first summation. The negative contributions to the second harmonic are represented by the second summation in equation (1) and for the propagations under consideration here can be approximated by the second term in that summation, $2u_1u_1^*$. Even this term, though, is negligible throughout most of the source to focal region propagation due to the relatively small amplitudes of the third harmonic. Thus, the nonlinear production of the second harmonic throughout most of the relevant propagation region is simply proportional to the square of the amplitude of the fundamental.

When n equals 3 in equation (1), the positive contributions to $\frac{\partial u_3}{\partial z}$ come from the first two terms comprising the first summation. These two terms sum to $3u_1u_2$. The first term of the corresponding negative contributions to the third harmonic is $3u_1u_1^*$. This term and its successors are negligible for all but the focal region of the highest intensity propagations considered herein.

The harmonic sources of the second harmonic and third harmonic are thus $1u_1u_1$ and $3u_1u_2$, respectively. At a given point in the field of the propagating beam then, the finite amplitude production of the second harmonic is proportional to the square of the fundamental harmonic's amplitude. The production of the second harmonic off the beam axis is very small since the amplitude of the fundamental beam there is quite small. The third harmonic is produced in proportion to the product of the first and second harmonics and thus its nonlinear production is even more strongly weighted towards the beam axis.

Also of interest is how the production of

-22-

the second and third harmonics vary with the z coordinate. Neglecting the effects of absorption and approximating the effects of focusing by assuming spherically-converging wave propagation, the amplitudes of the fundamental harmonic's mainbeam increases approximately linearly with distance from the source to the focus. At $z = F/2$ the on-axis amplitudes of the beam are about twice the corresponding source plane amplitudes. Following from the same approximations, the fundamental's mainbeam width at $z = F/2$ is about half its source plane width. Thus, at $z = F/2$ the area of the fundamental's mainbeam is about $1/4$ the corresponding source plane area. Together, this relation and the previous amplitude relation suggest that the second harmonic beam production rate versus z is constant ($z < F$), with rate losses due to diminishing fundamental mainbeam area balanced by the concurrent gains due to increased fundamental amplitudes.

The fact that the third harmonic production is proportional to the product of the fundamental and second harmonic amplitude, though, implies that the production of the third harmonic is strongly weighted towards the focal region. In Figure 10a the log-scaled, axial amplitudes of the fundamental, second, and third harmonics are displayed at 120a, 120b and 120c respectively. The source was the same focused 2 MHz Gaussian source considered above. The medium's parameters were again those of liver. Consistent with the above discussion, the amplitudes of the second harmonic exhibit a relatively large gain in its growth from low near field values to significant focal

-23-

amplitudes. The third harmonic exhibits an even higher gain, approximately duplicating the growth in gain from the fundamental to the second harmonic. Both harmonics, though, display post-focal region amplitude declines which parallel those of the fundamental.

In Figure 10b the corresponding log-scaled, focal plane ($z = 6$ cm, one way) radial beam profiles are displayed at 122a, 122b and 122c respectively. The fundamental profile drops 49.8 dB over the 1 cm radial range displayed. The second harmonic approximately squares this decline in dropping 85.3 dB. The third harmonic then continues the relationship in dropping 120.5 dB. These declines reflect the second and third harmonic, finite amplitude production rates discussed above. At $z = 8$ cm this relationship between the harmonic beam profiles continued to hold.

In Figure 11, the axial amplitudes of the 4 MHz second harmonic beam, shown at 124a, are overlaid with the corresponding 4 MHz fundamental harmonic beam, shown at 124b. The on-axis source of both of the respective propagations was 2 w/cm^2 . The axial curves have been normalized to be unity at $z = 6$ cm and log-scaled. The 4 MHz second harmonic focal amplitude was originally (pre-normalization) 15.0 db before the corresponding fundamental value. The two curves are close through the focal region and then depart shortly after $z = 6$ cm as the 4 MHz linear curve rapidly declines.

The results displayed in Figures 10a, 10b, and 11 show that the 4 MHz second harmonic beam may be

-24-

less susceptible to the defocusing effects of near field phase aberrations than a 4 MHz fundamental beam.

- 1 Since only a fraction of the second harmonic beam forms in the near field, only this fraction could be redirected or defocused by near field jitter. The corresponding 2 MHz fundamental beam, though, would
- 5 pass in its entirety through the aberration and suffer the consequent defocusing effects including increased focal plane sidelobe levels. Secondly, though, these higher 2 MHz sidelobe levels could, in turn, increase the off-axis nonlinear production of the 4
- 10 MHz second harmonic.

- In order to investigate the effects of tissue-based phase aberration on the characteristics of linear and nonlinear beams, planes of phase delay or jitter were introduced into linear and nonlinear
- 15 propagations of the focused Gaussian transducer. These phase delay planes were computed using measured time delays from 5 abdominal wall layers and 5 breast wall layers. The 5 abdominal wall specimens had layer thicknesses of 2.5, 2.0, 1.5, 1.5, and 1.0-3.0 cm (a
- 20 non-uniform layer) giving an average thickness of 1.9 cm. The 5 breast wall specimens had layer thicknesses of 1.5-2.5, 3.0-3.5, 3.5, 4.0 and 2.0-2.5 cm, giving an average thickness of 3.0 cm. The average thicknesses of the non-uniform layers were used to
- 25 compute the 5-layer averages. All of the measured abdominal wall and breast wall layers contained an outer skin layer.

- Each of the measured abdominal wall time delay planes was converted to an equivalent 2 MHz
- 30 phase delay plane. Each of these delay planes was

-25-

then scaled by 0.5 (i.e., each phase delay was reduced by a factor of 2) and then applied twice to a given beam propagation to represent the cumulative aspect of the actual tissue delays. In applying a single delay plane, the 2 MHz phase delay values were scaled for appropriate application to each harmonic present in the computed field. The first delay plane was encountered by the propagating field at $z = 0.5$ cm and the second plane at $z = 1.5$ cm. Further, subdivision c and subsequent applications of the abdominal wall delay data did not appear to be necessary since it did not significantly change the resulting focal plane fields. Thus, the 2 delay plane application scheme satisfied the thin lens approximation. The breast delay planes were likewise applied in two steps, the first at $z = 1$ cm and the second $z = 2$ cm. For both tissue types, the z placement was selected to represent the average slice thickness and also to be convenient for the Δz step size utilized by the linear propagation.

Figures 12a-12f depict pairs of one-way focal plane harmonic amplitude diameters and the corresponding average radii for an unjittered (or homogeneous) path and an abdominal wall-jittered propagation path. The propagation parameters of both mediums were again those of liver. In Figure 12a the corresponding focal plane diameters for the unjittered and jittered 2 MHz fields are overlaid at 132a and 132b respectively. In Figure 12b the corresponding average radii are shown at 134a and 134b respectively for the 2 MHz fields. The average radii were obtained by averaging the focal plane grid of amplitudes around

-26-

the axis. The corresponding results for the second harmonic 4 MHz are shown at 136a, 136b, 140a and 140b in Figures 12c and 12d, and for the fundamental 4 MHz fields are shown in at 142a, 142b, 144a, and 144b in Figures 12e and 12f. In all of these Figures, the unjittered field is displayed using solid curves.

In Figures 12a-12f, two basic effects of jitter are visible. The first is the increased sidelobe levels associated with the defocusing of the abdominal wall phase delays. This sidelobe effect is more prominent for both of the 4 MHz fields than for the 2 MHz fundamental field. The shorter wavelengths of these 4 MHz fields allows for greater de-focusing by the phase screen. The second harmonic 4 MHz field also gets a sidelobe level increase from the corresponding increase in the 2 MHz field's sidelobe levels.

The second impact revealed in Figures 12a-12f is that there are small changes in the mainlobes of all three harmonic fields. In particular, in the diameter Figures 12a, 12c and 12e, a shift in the peak or center of the jittered lobes can be seen. In the corresponding average Figures 12b, 12d and 12f though, the impact of the jitter is negligible down to approximately 20 dB below the peak on-axis value. Thus, the impact of this abdominal wall-jittering does not appear to involve any significant broadening of the mainlobes. Finally, both of the jittered 4 MHz mainlobes show decreases in peak amplitudes which are consistent with the increased energy present in the sidelobe regions.

Figures 13a-13d show two-way average radii

-27-

results for the same abdominal wall-jittered propagation. Each of the curves shown was obtained by radially averaging the corresponding two-way planar data and then scaling the on-axis value to unity. The 2 MHz fundamental, 4 MHz second harmonic, and 6 MHz third harmonic average radii are shown at 146a, 146b and 146c in Figure 13a. The 2, 4 and 6 MHz fundamentals are shown at 146a, 150a and 150b in Figure 13b. The average two-way sidelobe levels of the second and third harmonic can be seen to be significantly lower than those of any of the fundamental harmonics. In Figure 13c, the two 4 MHz average two-way profiles are shown at 152a and 152b. The 4 MHz fundamental curve has a slightly narrower mainlobe (9.5% at 20 dB down from peak) and significantly higher sidelobe levels than the 4 MHz second harmonic profile. Likewise, Figure 13d shows the 6 MHz fundamental and 6 MHz third harmonic two-way profiles at 154a and 154b. The 6 MHz fundamental curve is 10.9% narrower at the -20 dB level and also shows higher sidelobe levels than the third harmonic curve.

The results shown in Figures 13a-13d show that the second and third harmonics maintain lower sidelobe levels than the corresponding fundamental harmonics in propagating through abdominal wall. In Figures 14a and 14b, radial results obtained by averaging across 5 abdominal wall-jittered propagations are shown. In Figure 14a the normalized radial averages from the 5 two-way planar amplitude data sets are depicted. The average 2 MHz result is overlaid with the corresponding 4 MHz second harmonic

-28-

and 4 MHz fundamental curves 156a, 156b and 156c respectively. The -20 dB width of the 4 MHz

1 fundamental profile 156c is 6.6% narrower than the 4 MHz second harmonic profile 156b. The -20 dB width for this and multiple propagation averaged results are given in Table 3 of Figure 15.

5 Each of the two-way average profiles shown in Figure 14a was then radially summed. Each of the average profiles consisted of 499 radial position-magnitude value pairs, (r_i, m_i) , $i = 1, \dots, 499$. The 499 pairs discretely described the averaged of 5 abdominal wall-jittered two-way profiles over a radial extent of 1.2 cm. The discrete radial summation of the average profiles was defined by 499 radial position-summation value (r_j, s_j) pairs, where each s_j was defined by

$$15 \quad s_j = \pi \sum_{i=1}^j [(r_i^2 - r_{i-1}^2) \times (m_i + m_{i-1}) / 2]$$

The first term in the summation involves the (r_0, m_0) on-axis magnitude value.

20 The resulting radial summation or integration profiles are shown in Figure 14b. In Figure 14b, the integration profiles for the 2 MHz fundamentals, the 4 MHz second harmonics and the 4 MHz 25 fundamentals are shown at 160a, 160b and 160c respectively. Each of these integrated two-way profiles was scaled such that the value at an off-axis radial distance of 1.2 cm was unity. The elevated sidelobes of the 2 and 4 MHz fundamental profiles 30 cause their summation profiles to rise significantly

-29-

beyond the radial extent of the mainlobe. This additional rise represents the potential for scattering from the sidelobes to significantly reduce the contrast resolution of the image. In Table 3 the radial extent at which these integration profiles reach the 0.9 level is given. This radial extent is a measure of the sidelobe's potential to reduce the contrast resolution of an image. In this case, the second harmonic's radial extent is 38% less than that of its 2 MHz fundamental and 63% less than the corresponding 4 MHz fundamental radial extent.

Alternatively, the summation profiles depicted in Figure 14b offer the percent of the two-way field's amplitude inside or outside a given radius. For example, 91.7% of the 2 MHz fundamental's amplitude, 96.7% of the 4 MHz second harmonic's amplitude, and 88.6% of the 4 MHz fundamental's amplitude fall inside a radius of 0.25 cm. The corresponding percentages falling outside of 0.25 cm are 8.3%, 3.3% and 11.4%, respectively. Ratios of these outside percentages could be useful for inferring the relative contrasts offered in imaging a low scattering or void region of a given size. For a void region approximately 0.5 cm across, the 4 MHz second harmonic of this device might then provide 2.5 times (8 dB) higher contrast than the 2 MHz fundamental and 3.5 times (11 dB) higher contrast than the corresponding 4 MHz fundamental.

The previous two sets (at 2 and 4 MHz) of 5 abdominal wall jittered propagations were repeated at twice the source frequencies. Figures 15a and 15b depict the corresponding averaged results from these 4

-30-

MHz nonlinear and 8 MHz linear propagations. In Figure 15a the radial average amplitude curves obtained by averaging the 5 two-way data sets are shown. Figure 15a shows that the 8 MHz fundamental mainbeam 162c is broader than the corresponding 4 MHz fundamental mainbeam 162a. The jitter-imposed lateral resolution limits have been encountered and in fact exceeded at this point. Also note that the 8 MHz second harmonic mainbeam 162b is narrower than either of the fundamentals 162a and 162c. Figure 15b depicts the additional sidelobe corruption of the fundamental beams and the corresponding increase in the second harmonic's relative contrast resolution potential.

Figures 16a, and 16b show average results from 5 propagations through breast wall delay data. The normalized two-way average radii for the 2 MHz fundamentals, the 4 MHz second harmonics and the 4 MHz fundamentals are shown at 166a, 166b and 166c respectively in Figure 16a; and the corresponding radially integrated magnitudes are shown at 170a, 170b and 170c respectively in Figure 16b.

In Figures 16a and 16b the results considered were from 2 MHz nonlinear and 4 MHz linear propagations. In Figure 16a, the 4 MHz mainbeams 166c are significantly broadened over the corresponding abdominal wall-jittered mainbeams in Figure 14a. The sidelobe levels in Figure 16a are also higher than those in Figure 14a. These Figures show that the breast wall layers produced significantly more distortion than the abdominal wall layers. The second harmonic 4 MHz profile 166b is 8.4% narrower at the -20 dB level than the fundamental 4 MHz 166c. In

-31-

Figure 16b the integrated profile of the second harmonic 170b has a radial extent at the 0.9 level which is 48% less than the 2 MHz fundamental's 170a and 70% less than the 4 MHz fundamental's 170c. Thus, in the more distorting breast wall-jittered propagations, the relative advantages of the 4 MHz second harmonic were larger than in the abdominal wall propagations and included a slightly narrower mainbeam. The 4 and 8 MHz results for breast wall-jittered propagations follow closely the developments seen in the 4 and 8 MHz curves of Figures 16a and 16b. Results from these propagations are included in Table 3 of Figure 17.

Finally, in all of the jittered propagations considered, the second harmonic mainbeam was narrower than the fundamental mainbeam. The limits on the lateral resolution of the linear harmonics eventually put limits on the second and other higher harmonics, though.

The above discussions show the liver-path beam patterns for a focused Gaussian-apodized transducer operating at 2, 4 and 8 MHz. The non-phase aberrated propagations show that the second and higher harmonics formed through finite amplitude distortion have much lower sidelobe levels than their fundamental harmonic or the corresponding linear fundamentals. The finite amplitude production of these higher harmonic beams allow this sidelobe relationship to hold for any focused or unfocused transducer. Pulse propagation analysis shows that the higher harmonics formed in a propagating pulse-beam can be very well described by considering the harmonics produced in the

-32-

corresponding continuous wave propagation. Modeling results also show that second harmonic levels

1 sufficient for imaging purposes can be easily obtained within the field amplitude limits of the mechanical index.

5 The introduction of phase jitter as computed from measured propagation delays from slices of abdominal wall and breast wall causes the sidelobe levels of the second harmonic and fundamental beams to rise significantly. In all of the aberrated propagations considered, the two-way profile of the
10 second harmonic offered narrower -20 dB mainlobe widths and lower sidelobe levels than the fundamental beam which produced it. These same second harmonic profiles had slightly broader mainlobes at 4 MHz in abdominal wall-jitter propagations than the 4 MHz
15 fundamental profiles. In all other jittered propagations considered, though, the second harmonic offered slightly narrower mainlobes than the same-frequency fundamental and substantially lower sidelobe levels. Thus, second harmonic-based ultrasonic images
20 offer significant improvement in the lateral component of contrast resolution.

One obstacle to obtaining such images are artifacts from source contributions to the higher harmonic bandwidths. Figure 18a depicts such an
25 imperfect source wave. This on-axis waveform is depicted in particle velocity units and corresponds to a peak pressure of one half of that of the pulse shown in Figure 6a. The pulse in Figure 18a has the same form as that in Figure 6a with the exception of the
30 initial zero portion. In Figure 18b, the

-33-

corresponding source pulse spectrum is shown.

Significant energy content outside the 2 MHz bandwidth
1 is visible. The resulting computed focal waveform's
spectrum is depicted in Figure 18c and can be compared
to the earlier focal spectrum in Figure 6b. The log-
scaled depiction in Figure 18c shows that a simple
5 high pass filtering of the received spectrum contains
significant contributions from the source.

In Figure 18d, the resulting nonlinear
distortion pulse obtained by high pass filtering the
spectrum of Figure 18c is shown. This pulse can be
10 compared to the corresponding distortion pulse shown
in Figure 6c. In both cases, the cut-off frequency
was 3 MHz. In Figure 18d the full 8 microsecond
period of the computed pulse is shown so that the
ringing associated with the linear or source content
15 within the second harmonic bandwidth can be seen.
This ringing could adversely affect the axial
component of the contrast resolution and the lateral
gains associated with the lower sidelobe levels. In
this case, improvements can be obtained by increasing
20 the source amplitude so as to boost the second
harmonic bandwidth levels.

A two pulse scheme may be used to alleviate
or to eliminate such problems. In this method, two
source pulses are sent in place of a single pulse in
25 the image formation cycle. The two pulses are
identical in form but one is significantly lower in
amplitude. The received echo from this lower
amplitude pulse is then used to remove the linear
content from the high amplitude pulse. This is
30 accomplished by subtracting an appropriately-scaled

-34-

version of the received low amplitude signal from the corresponding high amplitude signal. The resulting
1 difference signal may then be high-pass filtered
followed by the normal sequence of image formation
steps. The high-pass filtering is preferred since
pulse analysis has revealed that the low frequency
5 content of the difference signal is radially wide-
spread and reduces the higher harmonic sidelobe
advantage.

Figures 19a, 19b and 19c show results
relevant to an implementation of the two pulse scheme
10 using the pulse shown in Figure 19a as the high
amplitude source pulse. In Figure 19a, the focal
spectrum from Figure 18c is shown at 176a and is
overlaid with the focal spectrum 176b produced by a
half-amplitude version of the same source. The low or
15 half amplitude spectrum was multiplied by two and
subtracted from the high amplitude spectrum to obtain
the difference spectrum depicted in Figure 19b. Note
that if the propagations had not involved finite
amplitude distortion, then this difference spectrum
20 would have been all zero. The difference spectrum was
then high pass filtered and inverse transformed to
obtain the effective on-axis distortion imaging pulse
shown in Figure 19c. The transition for the high pass
filtering was at 2.75 MHz. This same frequency was
25 also appropriate for filtering off-axis difference
spectra.

This two pulse scheme appears to be capable
of extracting the desired largely-second harmonic
images from any realistic ultrasonic imaging pulse.
30 The third harmonic bandwidth depicted in Figure 19b,

-35-

though, does not appear to be separable from the second harmonic bandwidth. The nodal depth between these harmonics is not deep enough. Thus, third harmonic (or largely-third harmonic) images do not appear to be easily obtainable with this two pulse scheme. Also, a two pulse scheme with a 1/4000 second interval between the respective high and low amplitude source pulses, does not have significant artifacts due to tissue or transducer motion.

The use of the second harmonic (plus a small contribution from higher harmonics) to form an image is an independent alternative to the phase correction-based schemes which have been and are being examined by other investigators as a means for improving the contrast performance of biomedical ultrasonic imaging.

The higher harmonics also offer additional opportunities for correcting for beam distortion. The amplitude of the third harmonics in the focal region is strongly affected by the fundamental's amplitude. As discussed above, the nonlinear production of the third harmonic is proportional to the product of the amplitudes of the fundamental and second harmonics. This means that much of the third harmonic's production occurs in the focal region where beam distortion can reduce the amplitude of the fundamental and second harmonic. Thus, an iterative scheme may be used to correct for beam defocusing using the amplitude of the received third harmonic for feedback.

Second harmonic images also provide for a means of reducing speckle. In particular, an image formed as the sum of a second harmonic image and the corresponding fundamental image would have less

-36-

speckle than either of the constituent images. Since
the second harmonic is twice the frequency of the
1 fundamental and has a largely constant phase
relationship with the fundamental in the mainlobe, the
second harmonics image's speckle pattern would be
conveniently out of phase with that of the
5 fundamental.

While it is apparent that the invention
herein disclosed is well calculated to fulfill the
objects previously stated, it will be appreciated that
numerous modifications and embodiments may be devised
10 by those skilled in the art, and it is intended that
the appended claims cover all such modifications and
embodiments as fall within the true spirit and scope
of the present invention.

15

20

25

30

35

-37-

CLAIMS

1. A method of imaging a sample comprising the
1 steps of:
generating an ultrasonic signal;
directing the ultrasonic signal into a sample,
wherein the sample reflects the signal;
5 receiving the signal reflected by said sample,
which received signal is distorted and contains a
first order and higher order component signals at
first and higher frequencies respectively;
forming an image from one of said higher order
10 component signals of the received distorted signal,
including the step of removing from the received
distorted signal the first order component thereof;
and
displaying said formed image.
15
2. A method according to Claim 1, wherein the
removing step includes the step of high-pass filtering
the received, reflected distorted signal to remove
therefrom the first order component thereof.
20
3. A method according to Claim 1, wherein:
the generating signal includes the steps of
generating first and second ultrasonic signals;
the directing step includes the steps of
25 directing the first and second ultrasonic signals into
the sample;
the receiving step includes the step of receiving
any first and second signals reflected and distorted
by said sample;
30 the forming step includes the steps of

-38-

- i) subtracting the received second distorted signal from the received first distorted signal to produce a resultant signal, and
- ii) forming the image from said resultant signal.

4. A method according to Claim 3, wherein the first and second signals are identical except that one is scaled up in magnitude by a factor x (greater than 1) relative to the other and the second signal is transmitted after the reception of the distorted first signal.

10

5. A method according to Claim 4, wherein the sample (i) distorts the first ultrasonic signal to produce a first distorted signal, (ii) reflects the first distorted signal, (iii) distorts the second ultrasonic signal to produce a second distorted signal, and (iv) reflects the second distorted signal.

6. A method according to Claim 5, wherein the forming step includes the step of:

scaling the smaller received distorted signal (corresponding to the unscaled transmitted signal) by the previously used scale factor x ;

next subtracting this scaled signal to produce a difference signal essentially without frequency content in the original transmitted bandwidth; and

forming the image from one of said higher order component signals of the difference signal.

7. A method of imaging a sample, comprising the steps of:

-39-

generating an ultrasonic signal:
directing the ultrasonic signal into a sample,
1 wherein the sample reflects the signal;
receiving the signal reflected by said sample,
which received signal is distorted and contains a
first order and higher order component signals at
5 first and higher frequencies respectively;
forming an image from one of said higher order
component signals of the received distorted signal;
and
displaying said formed image;
10 wherein said higher order component signals
include a second order component, and the forming step
includes the step of forming the image from said
second order component.

15 8. A method of imaging a sample, comprising the
steps of:
generating an ultrasonic signal:
directing the ultrasonic signal into a sample,
wherein the sample reflects the signal;
20 receiving the signal reflected by said sample,
which received signal is distorted and contains a
first order and higher order component signals at
first and higher frequencies respectively;
forming an image from one of said higher order
25 component signals of the received distorted signal;
and
displaying said formed image;
wherein the directing step includes the step of
maintaining the sample substantially free of any
30 contrast agent not naturally present in the sample.

-40-

9. A system for imaging a sample, comprising:
means for generating an ultrasonic signal;
1 means for directing the ultrasonic signal into a
sample, wherein the sample reflects the signal;
means for receiving the signal reflected by said
sample, which received signal is distorted and
5 contains a first order and higher order component
signals at first and higher frequencies respectively;
means for forming an image from one of said
higher order component signals of the received
distorted signal, said means for forming the image
10 including means for removing from the received
distorted signal the first order component thereof;
and
means for displaying said formed image;

15 10. A system according to Claim 9, wherein the
means for removing the first order component from the
received distorted signal includes a high-pass filter
to filter the received, reflected distorted signal to
remove therefrom the first order component thereof.

20 11. A system according to Claim 9, wherein:
the means for generating the ultrasonic signal
includes means for generating first and second
ultrasonic signals;
25 the means for directing the ultrasonic signal
into the sample includes means for directing the first
and second ultrasonic signals into the sample;
the receiving means includes means for receiving
any first and second signals reflected and distorted
30 by said sample;

-41-

the means for forming the image includes

- i) means for subtracting the received second
1 distorted signal from the received first distorted
signal to produce a resultant signal, and
ii) means for forming the image from said
resultant signal.

5

12. A system according to Claim 11, wherein the
first and second signals are identical except that one
is scaled up in magnitude by a factor x (greater than
1) relative to the other and the second signal is
10 transmitted after the reception of the distorted first
signal.

13. A system according to Claim 12, wherein the
sample (i) distorts the first ultrasonic signal to
15 produce a first distorted signal, (ii) reflects the
first distorted signal, (iii) distorts the second
ultrasonic signal to produce a second distorted
signal, and (iv) reflects the second distorted signal.

- 20 14. A system according to Claim 12, wherein the
forming means includes:

- means for scaling the smaller received distorted
signal (corresponding to the unscaled transmitted
signal) by the previously used scale factor x ; and
25 next for subtracting this scaled signal to
produce a difference signal essentially without
frequency content in the original transmitted
bandwidth; and

- means for forming the image from one of said
30 higher order component signals of the difference

35

-42-

signal.

- 1 15. A system for imaging a sample, comprising:
 means for generating an ultrasonic signal;
 means for directing the ultrasonic signal into a
sample, wherein the sample reflects the signal;
5 means for receiving the signal reflected by said
sample, which received signal is distorted and
contains a first order and higher order component
signals at first and higher frequencies respectively;
 means for forming an image from one of said
10 higher order component signals of the received
distorted signal; wherein said higher order component
signals include a second order component, and the
means for forming the image includes means for forming
the image from said second order component; and
15 means for displaying said formed image.

16. A system for imaging a sample, comprising:
 means for generating an ultrasonic signal;
 means for directing the ultrasonic signal into a
20 sample, wherein the sample reflects the signal;
 means for receiving the signal reflected by said
sample, which received signal is distorted and
contains a first order and higher order component
signals at first and higher frequencies respectively;
25 means for forming an image from one of said
higher order component signals of the received
distorted signal; and
 means for displaying said formed image;
 wherein the sample is substantially free of any
30 contrast agent not naturally present in the sample.

-43-

17. A method of imaging a sample, comprising the steps of:

- 1 generating a series of ultrasonic pulse signals;
 directing the ultrasonic pulse signals into a
sample, wherein the sample reflects the pulse signals;
 receiving the pulse signals reflected by said
5 sample, which received pulse signals are distorted and
contain a first order and higher order component
signals at first and higher frequencies respectively;
 forming an image from one of said higher order
component signals of the received distorted pulse
10 signals; and
 displaying said formed image.

18. A method of imaging a biological sample,
comprising the steps of:

- 15 generating an ultrasonic signal;
 directing the ultrasonic signal into the
biological sample, wherein the sample reflects the
signal;
 receiving the signal reflected by said sample,
20 which received signal is distorted and contains a
first order and higher order component signals at
first and higher frequencies respectively;
 forming an image from one of said higher order
component signals of the received distorted signal;
25 and
 displaying said formed image.

19. A system for imaging a sample, comprising:
 means for generating a series of ultrasonic pulse
signals;
30 means for directing the ultrasonic pulse signals

-44-

into a sample, wherein the sample reflects the pulse signals;

1 means for receiving the pulse signals reflected by said sample, which received pulse signals are distorted and contain a first order and higher order component signals at first and higher frequencies
5 respectively;

means for forming an image from one of said higher order component signals of the received distorted pulse signals; and

means for displaying said formed image.

10

20. A system for imaging a sample, comprising:

means for generating an ultrasonic signal;

means for directing the ultrasonic signal into a sample, wherein the sample reflects the signal;

15 means for receiving the signal reflected by said sample, which received signal is distorted and contains a first order and higher order component signals at first and higher frequencies respectively;

20 means for forming an image from one of said higher order component signals of the received distorted signal; and

means for displaying said formed image;

wherein the sample is a biological sample.

25 21. A method of imaging a sample, comprising the steps of:

generating an ultrasonic signal;

directing the ultrasonic signal into the biological sample, wherein the sample linearly reflects the signal;

30 receiving the signal linearly reflected by said

35

-45-

sample, which received signal is distorted and
contains a first order and higher order component
1 signals at first and higher frequencies respectively;
forming an image from one of said higher order
component signals of the received distorted signal;
and
5 displaying said formed image.

22. A system for imaging a sample, comprising:
means for generating an ultrasonic signal;
means for directing the ultrasonic signal into a
10 sample, wherein the sample linearly reflects the
signal;
means for receiving the signal linearly reflected
15 by said sample, which received signal is distorted and
contains a first order and higher order component
signals at first and higher frequencies respectively;
means for forming an image from one of said
20 higher order component signals of the received
linearly reflected, distorted signal; and
means for displaying said formed image.


25

30

35

1 / 2 5

FIG. 1



2 / 25

FIG. 2A






FIG. 2B




3 / 25

FIG. 2C



4 / 2 5

FIG. 3



5 / 2 5

FIG. 4A





FIG. 4B



6 / 2 5

FIG. 5A





FIG. 5B






7 / 25
FIG. 6A

FIG. 6B




8 / 25

FIG. 6C



9 / 2 5

FIG. 7



10 / 25

FIG. 8

On-axis source intensity (RMS w/cm ²)	Received 2nd harmonic level (dB)	Received 3rd harmonic level (dB)	Minimum pressure at the geometric focus (MPa)	Mechanical index (min. press./sqrt(f))
0.5	26.54	50.88	-0.45	0.32
1	23.64	45.02	-0.61	0.43
2	20.73	39.23	-0.84	0.59
4	17.94	33.66	-1.13	0.80
8	15.37	28.59	-1.50	1.06

FIG. 9

Focal length (cm)	2nd harmonic level at the focus (dB)	Received 2nd harmonic level (dB)
4	15.99	19.75
6	15.11	20.73
8	15.01	22.53
10	15.29	24.69
12	15.83	27.07




11/25
FIG. 10A

FIG. 10B



12 / 25

FIG. 11



13 / 25

FIG. 12A





FIG. 12B



14 / 25

FIG. 12C





FIG. 12D



15 / 25

FIG. 12E





FIG. 12F



16 / 25

FIG.13A





FIG.13B



17/25

FIG.13C





FIG.13D



18/25

FIG.14A





FIG.14B



19/25

FIG.15A





FIG.15B



20/25

FIG.16A





FIG.16B



21 / 25

FIG.17

Frequency / Medium	-20 dB full-width(cm)	0.9 Integrated full-width(cm)
2 MHz / ab wall	0.388	0.458
4 MHz / ab wall	0.242	0.542
2nd har 4 MHz / ab wall	0.258	0.332
8 MHz / ab wall	0.304	0.602
2nd har 8 MHz / ab wall	0.216	0.338
2 MHz / br wall	0.406	0.746
4 MHz / br wall	0.334	0.858
2nd har 4 MHz / br wall	0.308	0.504
8 MHz / br wall	0.462	0.880
2nd har 8 MHz / br wall	0.298	0.526

22 / 25

FIG. 18A





FIG. 18B



23 / 25

FIG. 18C





FIG. 18D



24 / 25

FIG.19A






FIG.19B



25 / 25

FIG. 19C



INTERNATIONAL SEARCH REPORT

information on patent family members

International Application No.

PCT/US 97/20162

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 5410516 A	25-04-95	DE 3829999 A	15-03-90
		AU 637059 B	20-05-93
		AU 4182489 A	02-04-90
		CA 1338175 A	19-03-96
		WO 9002517 A	22-03-90
		DK 31691 A	22-02-91
		EP 0357164 A	07-03-90
		JP 2677888 B	17-11-97
		JP 4501518 T	19-03-92
US 5577505 A	26-11-96	DE 19642379 A	07-08-97
		JP 9224939 A	02-09-97